



**Republika ng Pilipinas**  
 Republic of the Philippines  
**Kagawaran ng Tanggulang Pambansa**  
 Department of National Defense  
**DND BIDS AND AWARDS COMMITTEE**

July 20, 2017

**SUPPLEMENTAL BID BULLETIN NR. DND/AFP-ME-CCL-R-17-04**  
 (AFPMC Medical Equipment Acquisition Project)  
 (ACQUISITION OF ONE (1) LOT CARDIAC CATHETERIZATION  
 LABORATORY SYSTEM FOR THE ARMED FORCES OF THE  
 PHILIPPINES HEALTH SERVICE COMMAND (AFPHSC))

This Supplemental Bid Bulletin is issued to inform prospective bidders of the clarifications, modifications, amendments, or additional requirements that form part of the Bidding Documents for the AFPHSC Medical Equipment Acquisition Project.

**A. A portion in the Section VII. Technical Specifications, is hereby amended as follows:**

FROM	TO
III. Patient Table	III. Patient Table
1. Length: 285 - 300 cm	1. Length: 285 - 320 cm

**B. There will be changes in the Checklist of Eligibility and Technical Documents during Submission and Opening of Bid and Evaluation (SOBE) as follows:**

**FROM**

I. LEGAL DOCUMENTS: XXX	
II. TECHNICAL DOCUMENTS	
TAB	TECHNICAL DOCUMENTS
XXX	XXX
I	I1 Current and Valid Certificate of Manufacturer's Declaration of Conformity with:  a.ISO 13485:2003 Medical Devices: Quality Management System – Requirements for Regulatory Purposes.  b.International Electrotechnical Commission (IEC) standard or its equivalent national standard.  I2 Certification from the bidder that he/she is the Exclusive Distributor/Manufacturer of the said equipment in the Philippines.

	<p>13 Warranty Certificate from the Exclusive Distributor/Manufacturer in accordance with the terms specified in <b>Section VI. Technical Specifications.</b></p> <p>14 Certification that the bidder will provide trainings in accordance with the terms specified in <b>Section VI. Technical Specifications.</b></p> <p>15 Certification that the equipment is a brand new, unused and of the latest model (state of the art) or was never listed in any product recall worldwide.  "Brand new" means that the item to be delivered is manufactured within twelve (12) months from the opening of bids.</p> <p>16 Certification that the bidder must have a Cardiac-Catheterization Laboratory (Cath Lab) Machine repair facility in the Philippines and the repair of equipment shall be done in the Philippines.</p> <p>17 Certification from the bidder guaranteeing the availability of qualified service technician and/or bio-mechanical engineer for Cathlab within 48 hours (except weekends and holidays) upon receipt of notice from the AFPMC in case of equipment breakdown during the warranty period, and to further certify that in case the equipment will be pulled-out for repair or maintenance, they will provide a Service Unit that the end-user can use.</p> <p>18 Certification that the bidder will provide free quarterly preventive maintenance during warranty period with attached list of engineers or technical personnel who shall perform preventive maintenance of the product.</p> <p>19 Certification from the bidder guaranteeing the availability of maintenance services after the warranty period for a minimum period of five (5) years.</p> <p>110 Certification that the 95% uptime of the product is guaranteed within the warranty period and that any accumulated downtime in excess of 5% shall be added to the warranty period.</p>
--	--

**TO**

I. LEGAL DOCUMENTS:	
xxx	
II. TECHNICAL DOCUMENTS	
<b>TAB</b>	<b>TECHNICAL DOCUMENTS</b>
xxx	xxx
I	11 Current and Valid Certificate of Manufacturer's Declaration of Conformity with: <ul style="list-style-type: none"> <li>a. ISO 13485:2003 Medical Devices: Quality Management System – Requirements for Regulatory Purposes.</li> <li>b. International Electrotechnical Commission (IEC) standard or its</li> </ul>

equivalent national standard.

12 Certification from the bidder that he/she is the Exclusive Distributor/Manufacturer of the said equipment in the Philippines.

13 Warranty Certificate from the Exclusive Distributor/Manufacturer in accordance with the terms specified in **Section VI. Technical Specifications.**

14 **Duly Notarized** Certification that the bidder will provide trainings in accordance with the terms specified in **Section VI. Technical Specifications.**

15 **Duly Notarized** Certification that the equipment is a brand new, unused and of the latest model (state of the art) or was never listed in any product recall worldwide.

"Brand new" means that the item to be delivered is manufactured within twelve (12) months from the opening of bids.

16 **Duly Notarized** Certification that the bidder must have a Cardiac-Catheterization Laboratory (Cath Lab) Machine repair facility in the Philippines and the repair of equipment shall be done in the Philippines.

17 **Duly Notarized** Certification from the bidder guaranteeing the availability of qualified service technician and/or bio-mechanical engineer for Cathlab within 48 hours (except weekends and holidays) upon receipt of notice from the AFPMC in case of equipment breakdown during the warranty period, and to further certify that in case the equipment will be pulled-out for repair or maintenance, they will provide a Service Unit that the end-user can use.

18 **Duly Notarized** Certification that the bidder will provide free quarterly preventive maintenance during warranty period with attached list of engineers or technical personnel who shall perform preventive maintenance of the product.


19 **Duly Notarized** Certification from the bidder guaranteeing the availability of maintenance services after the warranty period for a minimum period of five (5) years.

110 **Duly Notarized** Certification that the 95% uptime of the product is guaranteed within the warranty period and that any accumulated downtime in excess of 5% shall be added to the warranty period.

C. The schedule for the Submission and Opening of Bids will be amended as follows:

FROM	TO
TBAL	1:30 P.M. 01 August 2017

Be guided accordingly.

  
**ASEC JOSUE S. GAVERZA, JR.**  
Chairperson, DND Bids & Awards Committee