



Republika ng Pilipinas
 Republic of the Philippines
Kagawaran ng Tanggulang Pambansa
 Department of National Defense

February 09, 2015

Supplemental Bulletin Nr. DND/AFP-MDE-N-15-01
 (Capability Upgrade of Medical and Dental Equipment Project)

Here under are the changes to the Bidding Documents for this Project:



1.

Reference	Amendments / Inclusions
	The Negotiation Proper shall be held on: <p style="text-align: center;"><u>17 February 2015 at 10:00 am</u></p>
Section VII Technical Specifications	Section VII Technical Specifications
LINE ITEM 4 :SELF-RETAINING BRAIN RETRACTOR	LINE ITEM 4: SELF-RETAINING BRAIN RETRACTOR
EQUIPMENT-SPECIFIC CHARACTERISTICS -2 pcs Yasargil Flex Arm/1 pc Coupling Heads and Bases -Adapter/ OR Table/Ball Socket joint AFTER SALES SERVICE The proponent must have a repair facility in the Philippines and the repair of the equipment shall be done in the Philippines.	EQUIPMENT-SPECIFIC CHARACTERISTICS -2 pcs Flex Rotary Arm/1 pc Coupling Heads and Bases -Adapter/ OR Table/Ball Socket joint Adapter connected to OR Table/Adapter connected to Skull Clamp Ball Socket Joint AFTER SALES SERVICE The proponent must have a repair facility in the Philippines and the repair of the equipment shall be done in the Philippines. OR In case of major repairs that can only be done outside the Philippines, all expenses incidental to the repair shall be shouldered by the proponent during the warranty period.

LINE ITEM 5: MOBILE OPERATING ROOM SPOTLIGHT	LINE ITEM 5: MOBILE OPERATING ROOM SPOTLIGHT
TECHNICAL SPECIFICATIONS With electronic transformer/ voltage regulator per equipment requirement	TECHNICAL SPECIFICATIONS With electronic transformer/ voltage regulator per equipment requirement
LINE ITEM 6: JOBST INTERMITTENT COMPRESSION MACHINE	LINE ITEM 6: JOBST INTERMITTENT COMPRESSION MACHINE
Main Voltage: 230 VA	Main Voltage: 220 VA

2. Be guided accordingly.

For the DND Bids and Awards Committee:


ASEC. EFREN Q. FERNANDEZ 
Vice-Chairman, DND Bids and Awards Committee